

My PhilGEPS

My Organization

My Profile

Opportunities

Catalogue

Directory

About PhilGEPS

Pending Task

Notices

Award Notices

Monday, August 17, 2020 08:29 AM

Venus SolanoMNWD

» Log-out

Bid Notice Abstract

Detail Tracking Report

Request for Quotation (RFQ)

Reference Number7179610

Procuring EntityMETROPOLITAN NAGA WATER DISTRICT

TitleSupply and Delivery of Laboratory Incubator (PR 20-07-0301)

Area of DeliveryCamarines Sur

 Printable Version

Solicitation Number: PR 20-07-0301 Trade Agreement: Implementing Rules and Regulations Procurement Mode: Negotiated Procurement - Small Value Procurement (Sec. 53.9) Classification: Goods Category: Laboratory Supplies and Equipment Approved Budget for the Contract: PHP 235,888.00 Delivery Period: 30 Day/s Client Agency: Contact Person: Venus C. Solano Procurement Assistant A 40 J.Miranda Ave., Naga City Camarines Sur Philippines 4400 63-54-4737813 Ext.117 63-54-4739288 mnwdbac@yahoo.com	Status	Active
	Associated Components	1
	Bid Supplements	0
	Document Request List	0
	Date Published	17/08/2020
	Last Updated / Time	17/08/2020 12:00 AM
	Closing Date / Time	20/08/2020 5:00 PM

Description

1 Laboratory Incubator 1 unit
PN: 62000
Origin: Europe or US

Standard Features:
RS 485/422 interface
Microprocessor controller with program functions;
Protection against temperature above or below the set value. Temperatures can be set with any accuracy of 0.1°C
Protein-free alarm output
Acoustic and visual alarms
Temperature selection monitor class 3.1
Bushing on rear panel
Control panel and door lockable simultaneously
Interior of high-quality stainless steel (W-St. 14301 ASTM 304)
One HTS tray (with Telescopic rails)
Factory certificate based on the Hettich QM System, mean value of 9 measurement points as per DIN 12880:2007-05 at 37°C

Technical Specifications:
Temperature range for incubators 1: 1K above ambient temperature up to + 65°C
Exterior dimensions (without bushing and door handle)
W X D X H in mm: 710 x 825 x 970
Interior dimensions W x D x H in mm: 535 x 690 x 420
Internal volume in liters: 150
Footprint in m2: 0.5
Weight in kg: 88
Number of trays provided as standard: 2 (1 standard + 1 HTS)

Temperature Fluctuation at +37°C : ± 0.1K
Temperature uniformity at +37°C : ± 0.2K
Temperature uniformity at + 25°C : ± 0.1K
Recovery Time after door has been opened for 30 s at + 37°C: 3 min
Energy consumption at 37°C: 0.025kWh
Noise level: 41 dB(A)
Power supply: 220-240 V 1 ~/50-60 Hz
Weight in kg: 98

Note:
1. VAT Inclusive
2. FOB MNWD, Naga City
3. Partial Delivery shall not be accepted
4. Supplier must submit Brochure